

## 510(k) Summary

K081804  
p. 1 of 2

This 510(k) Summary is being submitted in accordance with the requirements of 21 CFR 807.92.

<b>Submitter:</b>	Medtronic Vascular 3576 Unocal Place Santa Rosa, CA 95403 USA	<b>AUG - 5 2008</b>
<b>Contact Person:</b>	Catherine Priestley Regulatory Affairs Specialist Desk: 707.591.7205 Fax: 707.591.7138 catherine.priestley@medtronic.com	
<b>Date Prepared:</b>	25 June 2008	
<b>Trade Name:</b>	Pioneer Plus Catheter	
<b>Common Name:</b>	Diagnostic Ultrasound Transducer and Percutaneous Catheter	
<b>Classification Name:</b>	Diagnostic Ultrasound Transducer and Percutaneous Catheter	
<b>Predicate Device:</b>	Pioneer Catheter, K072155 (5 October 2007)	
<b>Device Description:</b>	The proposed device is a catheter which utilizes IVUS imaging and a hollow Nitinol needle to facilitate redirection and placement of a 0.014" OTW guidewire into peripheral vessels. The guidewire can then facilitate placement of subsequent devices. The device is a single use, sterile catheter.	
<b>Statement of Intended Use:</b>	The Pioneer Plus Catheter is intended to facilitate placement and positioning of catheters within the peripheral vasculature. The Pioneer Plus Catheter also provides an intraluminal cross-sectional ultrasound image of the area of interest to facilitate placement of guidewires beyond stenotic lesions (e.g., sub-total, total or chronic total occlusions) prior to additional intervention (i.e. PTCA, stent, etc.). The Pioneer Plus Catheter is not indicated for use in the coronary or cerebral vasculature.	
<b>Summary of Technological Characteristics:</b>	Both products have the same indications and substantially equivalent performance, function, and device characteristics.	

## 510(k) Summary

**Summary of Non-clinical  
Data:**

In vitro and in vivo testing were completed to assess substantial equivalence between the Proposed Pioneer Plus and Predicate Pioneer Catheters in terms of Indications, performance, function, device characteristics, materials, packaging, biocompatibility, sterilization, and stability.

**Conclusion from Data:**

The results of this testing demonstrate equivalence and support a determination of substantial equivalence between the Proposed and Predicate devices.



Food and Drug Administration  
10903 New Hampshire Avenue  
Document Control Center – WO66-G609  
Silver Spring, MD 20993-0002

Medtronic Vascular Inc.  
Ms. Catherine Priestley, CSSBB, CQE, CQA  
Regulatory Affairs Specialist  
3576 Unocal Place  
Santa Rosa, CA 95403

SEP 18 2013

Re: K081804  
Trade/Device Name: Pioneer Plus Catheter  
Regulation Number: 21 CFR 870.1250  
Regulation Name: Percutaneous Catheter  
Regulatory Class: Class II  
Product Code: PDU, ITX  
Dated: June 25, 2008  
Received: June 26, 2008

Dear Ms. Priestly:

This letter corrects our substantially equivalent letter of August 5, 2008.

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA).

You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration. Please note: CDRH does not evaluate information related to contract liability warranties. We remind you, however, that device labeling must be truthful and not misleading.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.

Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); medical device reporting (reporting of medical

device-related adverse events) (21 CFR 803); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638 2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). For questions regarding the reporting of adverse events under the MDR regulation (21 CFR Part 803), please go to <http://www.fda.gov/MedicalDevices/Safety/ReportaProblem/default.htm> for the CDRH's Office of Surveillance and Biometrics/Division of Postmarket Surveillance.

You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 796-7100 or at its Internet address <http://www.fda.gov/MedicalDevices/ResourcesforYou/Industry/default.htm>.

Sincerely yours,



for

Bram D. Zuckerman, M.D.  
Director  
Division of Cardiovascular Devices  
Office of Device Evaluation  
Center for Devices and Radiological Health

Enclosure



**Medtronic**

*Alleviating Pain-Restoring Health-Extending Life*

CONFIDENTIAL - May not be reproduced  
without written permission from Medtronic, Inc.

## Indications for Use

510(k) Number K081804

Device Name: Pioneer Plus Catheter


### Indications for Use:

"The Pioneer Plus Catheter is intended to facilitate placement and positioning of catheters within the peripheral vasculature. The Pioneer Plus Catheter also provides an intraluminal cross-sectional ultrasound image of the area of interest to facilitate placement of guidewires beyond stenotic lesions (e.g., sub-total, total or chronic total occlusions) prior to additional intervention (i.e. PTCA, stent, etc.). The Pioneer Plus Catheter is not indicated for use in the coronary or cerebral vasculature."

Prescription Use   X   AND/OR Over-The-Counter Use         
(Part 21 CFR 801 Subpart D) (21 CFR 807 Subpart C)

(PLEASE DO NOT WRITE BELOW THIS LINE-CONTINUE ON ANOTHER  
PAGE IF NEEDED)

Concurrence of CDRH, Office of Device Evaluation (ODE)

  
\_\_\_\_\_  
(Division Sign-Off)  
Division of Cardiovascular Devices  
510(k) Number K081804